

**FORM TO BE USED BY TSC FOR FINALISING PRE-BID MINUTES**

**File No. 14-VI/HM (1143)/23PB/T-175**

**Dates: 02/01/2024**

**TSC Minutes (To be typed clearly by the I/O)**

Based on the Pre-bid meeting and recommendation of I/O, following changes have been made in the specifications:

<b>Original Specifications</b>	<b>Final Specifications</b>
<b>Annexure -I</b>	<b>Annexure –I (No Revision)</b>

The file is forwarded to Purchase Section for uploading the final specifications and TSC minutes on the website and CPPP Portal.

Declaration: We hereby declare that we have no conflict of interest with any of the bidder in this tender

### Triple Point of Argon with Realization Apparatus

#### Specifications:

Fixed Point Temperature:  $-189.3442\text{ }^{\circ}\text{C} \pm 0.0008^{\circ}\text{C}$

Argon Gas Purity: 99.9999% or better

Argon Cell Immersion Depth: 120 mm or higher

Accuracy:  $\pm 0.0008\text{ }^{\circ}\text{C}$  or better

Entrant Well Inner Diameter ID: 7mm to 11 mm

Plateau Duration: Minimum 4 Hour

Power: As per Indian Standard

Calibration: As per ISO IEC 17025: 2017 or NMI Traceable Certificate

Warranty: One Year

Installation: Installation will be done by Supplier at CSIR-NPL.

Acceptance Criteria: All quoted capabilities shall be demonstrated at CSIR-NPL, New Delhi.